

This drawing includes confidential information which is the property of ASHCROFT INC. This copy is loaned for a specific purpose, with the understanding that it will not be reproduced or used for any other purpose or disclosed to others, and it will be returned on demand.

ASHCROFT INC.

250 EAST MAIN STREET
STRATFORD, CT 06614

NOTE 1: BRACKET POSITIONED FOR 3/16" PANEL
NOTE 2: PANEL THICKNESS RANGES FROM 1/16" TO 1/4"
NOTE3: PANEL CUTOUT : 2.66[67.5] ±.02[0.4]







TABLE 1
CONNECTION SIZE AND TYPE

| CODE | DESCRIPTION |
|------|-----------------------------------|
| 02 | 1/4 NPT (Shown) |
| 01 | 1/8 NPT |
| HH | 10-32/2B Brass Female Thread |
| HD | 1/8" I.D. Tubing Brass Hose Barb |
| HE | 3/16" I.D. Tubing Brass Hose Barb |
| HF | 1/4" I.D. Tubing Brass Hose Barb |

1/4 NPT (STANDARD)
(SEE TABLE 1 OPTIONS)



| REV. | ECO | DESCRIPTION | BY | DATE |
|------|---------|-----------------|-----|-----------|
| A | 0008335 | INITIAL RELEASE | RFF | 9/28/2015 |

DIMENSIONS IN [] ARE MILLIMETERS

| This print certified correct for | | General Dimensions | | | |
|----------------------------------|-----------|------------------------|---------------|----------|------|
| Customer's Name | | 25 1490 02B XUC SERIES | | | |
| Customer's Order No. | Drawn RFF | Date 9/28/15 | Date of Issue | | |
| Ashcroft Order No. | Checked | Date | SHEET-1 | | Rev. |
| Certified by | Date | Approved | Date | 070A4113 | A |